



## BASIC WATER PURIFICATION SYSTEM BBPS-201

## BASIC WATER PURIFICATION SYSTEM BBPS-201

This basic series is ideal for wide range of applications. It produces RO, Deionized water and Ultrapure water. The organic rejection rate is greater than 99% using reverse osmosis. The resistivity reaches up to 18.2M $\Omega$ .cm which completely meets the highest grade I standard.

Used in Laboratory, Manufacturing, Reefkeeping, Aquarium, Laboratory, Research.

Also known as Laboratory Ultrapure water system.

## BBPS-201 BASIC WATER PURIFICATION SYSTEM



Human engineering design, high-strength, streamline plastic shell.

One time injection molding process case, material: Polypropylene PP.

Elegant and compact case, integrating pre-filter, RO, DI, UV, UF and terminal filter into one.

All filters are built-in, for the smallest outside space.

Top cap of pre-filters in the case can be rapidly opened to replace the pre-filters without opening the case.

With electronic pressure sensor and microcomputer controlling, the system automatically produces pure water.

Automatic stop without water, automatic stop when water tank full, automatically cutting off water when pump stopping,

guaranteeing 24 hours' work.

Self-flushing of the reverse osmosis membrane, extend the life of RO membrane.

On-line resistivity monitor, with apheliotropic LCD display, to detect the quality of deionized or ultrapure water.

Attached portable TDS (total dissolved solid)/conductivity test pen, with dry cell design, to detect the quality of tap water and RO water.

Different external tanks (optional) to meet every need and assure ample water-supply.

Pretreatment cartridges, RO module, ultrapure cartridges, all designed to modularization independently. Easy to maintenance and replacement.

Pipeline and fast-plug adaptor with NSF authorization, assure high quality ultrapure water.

DOW's RO membrane, ensure stable operation and high desalinization rate.

4 ultrapure cartridges, with DOW's nuclear-grade polishing resin, ensure ultrapure water's quality up to 18.2 M $\Omega$ .cm, with the lowest TOC dissolution.

Double wavelength (185&254nm) ultraviolet lamp module, restrain bacteria's increase and reduce TOC.

MWCO 5000D ultrafiltration module, effectively eliminate endotoxin precise cell cultivating and IVF.

(0.45+0.1) $\mu$ m double layer PES terminal disinfection filter, assure the quality absolutely axenic.

## SPECIFICATIONS

Model	BBPS-201
Feed Water Requirements*	
Water Inlet	RO water, Distilled water, Deionized water
Temperature	5-45°C

Pressure	1atm*
Flow Procedure**	AC+DI+TF
Ion rejection rate	-
Organic rejection rate	-
Particles and bacteria rejection rate	-
Bacteria	<0.1 cfu/ml
Output(25°C)****	Utmost up to 2.0 L/min (less output with UF cartridge)
Pure water outlet	Deionized water and Ultrapure water
Water Quality Monitor	Portable TDS/conductivity test pen + on-line resistivity monitor
DimensionLxWxH	410x220x420 mm
Weight	20 kg
Standard configuration	Main body (Including 1 set of cartridges)+ accessory bag
Power Consumption (W)	72 W
Power Supply	AC110-220 V, 50/60 Hz
Note	<p>*The feed water quality will influence the pure water's quality and cartridges life-span.</p> <p>**AC:active carbon, DI:ion exchange, UV:ultraviolet, UF:ultrafiltration, TF:terminal microfiltration. ***Value of number will be influenced by feed water quality. ****The output will decrease with terminal filter or UF cartridge.</p>
Deionized water quality	
Resistivity	>5 MΩ.cm
Particle(>0.2µm)	<1/ml
Ultrapure Water Quality	
TOC***	<10 ppb
Heavy metal ion	<0.1 ppb



**Biolab Scientific Ltd.**

3660 Midland Avenue, Suite 300, Toronto, Ontario M1V 0B8, Canada

Email: [info@biolabscientific.com](mailto:info@biolabscientific.com) | Website: [www.biolabscientific.com](http://www.biolabscientific.com)