



LABORATORY WATER PURIFICATION SYSTEM BLPS-401

LABORATORY WATER PURIFICATION SYSTEM BLPS-401

Laboratory water purification system uses double stage reverse osmosis technology. It produces double stage RO water, Deionised, EDI and ultrapure water. These systems have 3 way on-line water quality sensor, multiple alarm with unique design and it has easy-to-replace cartridges pack unit.

Used in Laboratory, Manufacturing, Reefkeeping, Aquarium.

Also known as Laboratory Deionized water system.

BLPS-401 LABORATORY WATER PURIFICATION SYSTEM



With tap water inlet, to produce RO water and ultrapure water, quality can reach to above 10 MΩ.cm.

Built-in 20 liters airtight plastic pressure water tank

Built-in 13 liters high-capacity polishing resin cartridge

Unique design and easy-to-replace cartridges pack unit.

Data storage and RS 232/USB communication port.

3 way on-line water quality sensor, multiple alarm.

Life-span of cartridges' display and alarm.

System circulation function, system sterilization procedure. (optional)

Molding process, high-strength, streamline plastic shell.

The graphic display clearly indicates all system's parameters. From water quality to knowing when it is time to change the purification pack, you'll see at a glance what is needed.

For ease-of-use, the main purification technologies are contained in an innovative all-in-one pack that means you can change it in just a couple of minutes.

The system requires no special installation, connect the system to your tap water supply it's ready to use.

SPECIFICATIONS

Model	BLPS-401
Feed Water Requirements*	
Water Inlet	Tap water
Temperature	5-45°C
Pressure	1.0-4.0 Kg/cm ²
Bacteria	<0.1 cfu/ml
Dimension LxWxH	570x600x1500 mm
Weight	60 kg
Power Consumption (W)	120 W
Power Supply	AC110-220 V, 50/60 Hz
Note	*The quality of output water accords with the quality of inlet water.
Deionized water quality	
Resistivity	>10 MΩ.cm
Conductivity	-
Particle(>0.2µm)	<1/ml
Ultrapure Water Quality	
Heavy metal ion	<0.1 ppb
Feed Water Requirements	
Output	60 L/hrs

Conductivity of RO water quality	< tap waterx4%
----------------------------------	----------------



Biolab Scientific Ltd.

3660 Midland Avenue, Suite 300, Toronto, Ontario M1V 0B8, Canada

Email: info@biolabscientific.com | Website: www.biolabscientific.com