



## GLOSS METER BMET-906

## GLOSS METER BMET-906

A glossmeter measures the specular reflection (gloss) of a surface. It determines the gloss by projecting a beam of light onto a surface at a fixed angle and intensity then measures the amount of reflected light at an opposite but equal angle.

## BMET-906 GLOSS METER



## SPECIFICATIONS

Model	BMET-906
Measuring Angle	60°
Measuring Area (mm)	60°:9X15
Measuring Range	60°:0~1000 GU
Division Value	0.1 GU
Measuring Time	0.5 s
Repeatability	0~100GU:±0.2GU 100-1000GU:±0.2G U %
Accurate	Conform with JJG696 first grade requirements of gloss meter
Auto Power-off Time	Within 30 s-120 s
Long-time Calibration	Automatically finished calibration
Language	Chinese English
Storage	35000(Basic mode SStatistic mode :15000; Continuous mode : 10000; QC mode: 10000)
Display	2.3 inch black and white screen
Size	160x52x84 mm
Weight	About 300 g
Power Supply	1 pc dry-cell battery(can measure 10000 times) or use USB Charge
Interface	USB & Bluetooth
PC Software	GQC6 QC software for quality report and more extended functions
Operation Temperature Range	0~40 °C (32~104 °F)
Storage Temperature Range	- 20~50 °C (-4~122 °F)
Humidity Range	< 85% RH, without condensation
Standard Accessories	User manual, calibration plate,USB cable, user manual, QC software (Download official website) calibration plate,user manual
Optional Accessories	Miniature printer



**Biolab Scientific Ltd.**

3660 Midland Avenue, Suite 300, Toronto, Ontario M1V 0B8, Canada  
Email: [contact@biolabscientific.com](mailto:contact@biolabscientific.com) | Website: [www.biolabscientific.com](http://www.biolabscientific.com)