



CCS MACHINE (MANUAL SAMPLING) BPTL-109

CCS MACHINE (MANUAL SAMPLING) BPTL-109

Viscosity of an oil is defined as its resistance to flow and Viscometer measures the flow. Monitoring viscosity is one of the most important components of any oil analysis program. Minute changes in viscosity at operating temperatures can be magnified to the extent that an oil is no longer able to provide adequate lubrication and efficiency in preventing friction between moving parts.

Used in Petroleum Industry, Oil and Gas Industry.

BPTL-109 CCS MACHINE (MANUAL SAMPLING)



Automatically calculate calibration error of standard oil. Viscosity measuring range:1500-25000cp. Check the accuracy of data for users. Can edit dozens of standard oil.

SPECIFICATIONS

Model	BPTL-109
Test temperature(°C)	-5,-10 -15,-20,-25,-30,-35
Temperature controlling range	Can reach to $\pm 0.02^{\circ}\text{C}$
Viscosity measuring range	1500-25000cp
Cooling system	Peltier cooling system,can measure any temperature rapidly
Calibration	Automatically calculate calibration error of standard oil
Rotor	Upgraded rotor,low torque test status
System	Window operation system,XP,Win7



Biolab Scientific Ltd.

3660 Midland Avenue, Suite 300, Toronto, Ontario M1V 0B8, Canada

Email: info@biolabscientific.com | Website: www.biolabscientific.com