



Product Image Coming Soon

MODERATELY HIGH TEMPERATURE PISTON DEPOSITS APPARATUS BPTL-401

MODERATELY HIGH TEMPERATURE PISTON DEPOSITS APPARATUS BPTL-401

The instrument is used to test thermo-oxidation engine oil simulation. The procedure includes determination of mass of deposit formed on a specially constructed test rod after oxidative and catalytic conditions at 285 °C is repetitively exposed to passage of 8.5 g of engine oil.

Used in Petroleum Industry, Oil and Gas Industry.

BPTL-401 MODERATELY HIGH TEMPERATURE PISTON DEPOSITS APPARATUS



Product Image Coming Soon

User can set working time, automatic stop after finishing test. Catalyst magnetic mixing device can set mixing time and speed. All parameters can be set on touch screen, one-click start test.

SPECIFICATIONS

Model	BPTL-401
Gas flow	Electronic flow controller accurately controls the gas flow at $10\text{ml} \pm 0.2\text{ml}/\text{min}$
Mixing device	Catalyst magnetic mixing device can set mixing time and speed
Display	All parameters can be set on touch screen, one-click start test
Optional	ASTM D6335 and SH/T0750 test components



Biolab Scientific Ltd.

3660 Midland Avenue, Suite 300, Toronto, Ontario M1V 0B8, Canada
Email: contact@biolabscientific.com | Website: www.biolabscientific.com