



ILLUMINANCE METER

ILLUMINANCE METER BJU-4201

DIGITAL ILLUMINANCE METER



SPECIFICATIONS

Model	BJU-4201
Light Measure Range	0.1 to 400,000 Lux(Fc switchable)
Accuracy	±4%(0.1 to 10,000Lux);±10% above 10,000Lux
Resolution	1 Lux/0.1Fc
Sampling	0.5sec
Temperature Range	-10~50°C(14 to 122°F)
Temperature Accuracy	±1°C
Humidity Range	20%~90%RH
Humidity Accuracy	±5%
Operation Temperature	0 to 50°C(32 to122°F)
Operation Humidity	10 to 90%RH
Storage Temperature	-10 to 60°C(14 to140°F)
Storage Humidity	10 to 75%RH
Net Weight	0.35kg
Gross Weight	0.5kg
External Size (WxDxH)	55x165x35mm
Shipping Size(WxDxH)	100x220x50mm
Alt Name	Digital Illuminance Meter



Biolab Scientific Ltd.

Trillium Executive Center, East Tower, 675 Cochrane Dr, Markham, Ontario L3R 0B8, Canada

Email: info@biolabscientific.com | Website: www.biolabscientific.com