



ATOMIC ABSORPTION SPECTROPHOTOMETER BAAS-603

ATOMIC ABSORPTION SPECTROPHOTOMETER BAAS-603

ATOMIC ABSORPTION SPECTROPHOTOMETER

An analytical technique used to measure the concentrations of elements in a sample. The system is incredibly sensitive and can detect down to micrograms (?g). It is performed by focusing a beam of known wavelength of ultraviolet (UV) light through a flame and into a detector.



Six lamp flame/graphite furnace integrated machine. With automatic safety protection function, anti-tempering automatic gas path protection, acetylene gas leak- age alarm, automatic shutdown system, abnormal automatic power-off.

SPECIFICATIONS

Model	BAAS-603
Grating	1800 l/mm
Wavelength Range	190-900 nm
Spectral Bandwidth	0.1, 0.2, 0.4, 1.0, 2.0 nm (automatic adjustable)
Wavelength Accuracy	≤ 0.15 nm
Wavelength Repeatability	± 0.1 nm
Baseline Stability	≤ ±0.002 A /30 minutes (static), ≤ ±0.005 A /30 minutes (dynamic)
Light Source	≤ 6 lamps automatic turret, automatic alignment
Power	Double cathode power built-in high performance lamps
Flame atomizer	
- Detection Limits (Cu)	0.002 µg/mL
- Precision	RSD ≤ 0.5 %
- Combustion Head	Metal Titanium combustion head
- Atomizer	Efficient glass atomizer
- Atomizing Chamber	Explosion proof corrosion resistant material spray chamber
- Control System	Automatic PC control three light brick, automatic alignment, automatic optimization and automatic ignition
- Safety Protection	With automatic safety protection function, anti-tempering, automatic gas path protection, acetylene gas leakage alarm, automatic shutdown system, abnormal automatic power-off
- Background Correction	Deuterium background correction: correction of the 1A background
Data processing	
- Measurement Methods	Flame method, Hydride method
- Concentration Calculation Method	Standard curve method (1 - 3 times curve), automatic matching, the standard addition method
- Repetition Survey Frequency	1-99 times, calculating the average value, standard deviation and relative standard deviations are given
- Results Print	Parameters print, data and graphics print, export WORD and EXCEL document. Simple operation, lamp position rotating, automatic ignition through software
Communication Interface	Computer and USB interface communication
Power Requirements	220 V (+5 % ~ -10 %), 50/60 Hz; 5000 VA
Environment Temperature	+15 °C ~ +35 °C
Relative Humidity	20 ~ 80 %
Alt Name	Atomic Absorption Spectrophotometer



Biolab Scientific Ltd.

Trillium Executive Center, East Tower, 675 Cochrane Dr, Markham, Ontario L3R 0B8, Canada
Email: info@biolabscientific.com | Website: www.biolabscientific.com