



# LABORATORY LOW SPEED CENTRIFUGE BCBL-108

# LABORATORY LOW SPEED CENTRIFUGE BCBL-108

## TABLETOP LOW-SPEED CENTRIFUGE



It has micro-computer control, electronic lock and DC brushless motor which is low noise and high speed accuracy.

Digital LCD screen make visualized operation conveniently.

Up to 9 user programs storage meet more demands.

It has self-locking, over-speed detection, over-temperature detection, imbalance detection that is automatic alarm device.

It has 9 levels acceleration and 10 levels deceleration to be setting according to the different requirement. Slow down freely is available.

Silastic sealed ring comfort to GMP, US FDA certificated.

## SPECIFICATIONS

Model	BCBL-108
Max speed	6000r/min
Max RCF	5120xg
Max capacity	16x15ml / 24x10ml
Speed accuracy	±30r/min
Time setting	1min~99min59s
Noise	<65dB(A)
Power supply	AC110V/AC220V, 50Hz~60Hz
Consumption	250W
Dimension (LxWxH)	410mmx490mmx310mm
Packing size (LxWxH)	500mmx580mmx400mm
Net weight (No rotor)	25kg
Alt Name	Tabletop Low-Speed Centrifuge

## ACCESSORIES FOR PURCHASE

No	Name	Description
1	Angle Rotor	Capacity: 12x15 ml Max Speed: 6000 r/min / 5000 r/min Max RCF: 5120 xg / 4100 xg
2	Angle Rotor	Capacity: 24x10 ml Max Speed: 6000 r/min / 5000 r/min Max RCF: 5120 xg / 4100 xg
3	Swing rotor	Capacity: 16x15 ml Max Speed: 4200 r/min Max RCF: 2760 xg Rotor Angle: =90°
4	Adaptor	4x4x15 ml 1.5,5,7,10 ml
5	Swing rotor	Max Speed: 4200 r/min Max RCF: 2760 xg Rotor Angle: =60°
6	Adaptor	2x12x5 ml 5 ml / 7 ml
7	Adaptor	12x2x10 ml 0.5 ml, 1.5 ml, 5 ml / 7 ml
8	Adaptor	12x15 ml 0.5 ml, 5 ml, 10 ml, 5 ml / 7 ml

9	Swing rotor	Capacity: 4x50 ml Max Speed: 4200 r/min Max RCF: 2760 xg Rotor Angle: =60°
10	Adaptor	50 mlx4 4x2x1.5 ml 4x5 ml / 7 ml 4x10 ml 4x15 ml 4x20 ml



1



2



3



4



5



6



7



8



9



10



**Biolab Scientific Ltd.**

Trillium Executive Center, East Tower, 675 Cochrane Dr, Markham, Ontario L3R 0B8, Canada

Email: [info@biolabscientific.com](mailto:info@biolabscientific.com) | Website: [www.biolabscientific.com](http://www.biolabscientific.com)