



LABORATORY LOW SPEED CENTRIFUGE BCBL-112



LABORATORY LOW SPEED CENTRIFUGE BCBL-112

TABLETOP LOW-SPEED CENTRIFUGE



It has micro-computer control, electronic lock and DC brushless motor which is low noise and high speed accuracy.

Digital LCD screen make visualized operation conveniently.

Up to 9 user programs storage meet more demands.

It has self-locking, over-speed detection, over-temperature detection, imbalance detection that is automatic alarm device.

It has 9 levels acceleration and 10 levels deceleration to be setting according to the different requirement. Slow down freely is available.

Silastic sealed ring comfort to GMP, US FDA certificated.

SPECIFICATIONS

| | |
|-----------------------|-------------------------------|
| Model | BCBL-112 |
| Max speed | 5000r/min |
| Max RCF | 4100xg |
| Max capacity | |
| Speed accuracy | ±30r/min |
| Time setting | 1min~99min59s |
| Noise | <65dB(A) |
| Power supply | AC110V/AC220V, 50Hz~60Hz |
| Consumption | 250W |
| Dimension (LxWxH) | 410mmx490mmx310mm |
| Packing size (LxWxH) | 500mmx580mmx400mm |
| Net weight (No rotor) | 25kg |
| Alt Name | Tabletop Low-Speed Centrifuge |

ACCESSORIES FOR PURCHASE

| No | Name | Description |
|----|-------------|--|
| 1 | Angle Rotor | Capacity: 12x15 ml Max Speed: 6000 r/min / 5000 r/min Max RCF: 5120 xg / 4100 xg |
| 2 | Angle Rotor | Capacity: 24x10 ml Max Speed: 6000 r/min / 5000 r/min Max RCF: 5120 xg / 4100 xg |
| 3 | Swing rotor | Capacity: 16x15 ml Max Speed: 4200 r/min Max RCF: 2760 xg Rotor Angle: =90° |
| 4 | Adaptor | 4x4x15 ml 1.5,5,7,10 ml |
| 5 | Swing rotor | Max Speed: 4200 r/min Max RCF: 2760 xg Rotor Angle: =60° |
| 6 | Adaptor | 2x12x5 ml 5 ml / 7 ml |
| 7 | Adaptor | 12x2x10 ml 0.5 ml, 1.5 ml, 5 ml / 7 ml |
| 8 | Adaptor | 12x15 ml 0.5 ml, 5 ml, 10 ml, 5 ml / 7 ml |

| | | |
|----|-------------|---|
| 9 | Swing rotor | Capacity: 4x50 ml Max Speed: 4200 r/min Max RCF: 2760 xg Rotor Angle: =60° |
| 10 | Adaptor | 50 mlx4 4x2x1.5 ml 4x5 ml / 7 ml 4x10 ml 4x15 ml 4x20 ml |



1



2



3



4



5



6



7



8



9



10



Biolab Scientific Ltd.

Trillium Executive Center, East Tower, 675 Cochrane Dr, Markham, Ontario L3R 0B8, Canada
Email: info@biolabscientific.com | Website: www.biolabscientific.com