BioLAB

PRODUCT CATALOG

# BENCHTOP LOW SPEED CENTRIFUGE BCBL-304

CE ISO9001 CERTIFIED

www.biolabscientific.com

# LOW SPEED CENTRIFUGE BCBL-304

TABLETOP LOW SPEED CENTRIFUGE

1. Microprocessor control, less noisily, it is widely used to qualitative analysis of blood serum, plasma and urea in the fields of hospital, blood center, laboratory and biochemistry.
2. Brushless motor, free maintenance, no powder pollution, quick in speed up and down.
3. The flexible axle driven system which drive the rotor directly, smooth in operation and small vibration.
4. There are many rotors for your choice, suitable for different specifications meet customers' different requirements of separation.
5. Micro computer control system, LED display the RCF,time and speed.
6. Electric lid lock, compact design, super speed and imbalance protection.

## SPECIFICATIONS

| Model | BCBL-304 |
|---|---|
| Max. Speed | 5000rpm |
| Max. RCF | 4420xg |
| Max. Capacity | 6x100ml |
| Time Range | 0~99min59sec |
| Noise | ≤ 55dBA |
| Speed Accuracy | ± 20r/min |
| Voltage(V/Hz) | AC 220V/110V 50HZ/60HZ |
| Size (W x D x Hmm) | 540x370x280mm |
| Net Weight(Kg) | 32KG |
| Certificates | CE,ISO & Calibration report are available |
| Alt Name | Tabletop Low Speed Centrifuge |

## ACCESSORIES FOR PURCHASE

| No | Name | Max Speed (rpm) | Max Volume (ml) | Max. RCF (xg) |
|---|---|---|---|---|
| 1 | Swing out rotor | 4000 | 4x50ml/100ml | 2250 |
| 2 | Swing out rotor | 4000 | 16x15ml | 2150 |
| 3 | Swing out rotor | 4000 | 24x10ml | 2220 |
| 4 | Angle rotor | 4000 | 18x10ml | 2250 |

1

2

3

4

Biolab Scientific Ltd.

Trillium Executive Center, East Tower, 675 Cochrane Dr, Markham, Ontario L3R 0B8, Canada

Email: info@biolabscientific.com | Website: www.biolabscientific.com