



LABORATORY LOW SPEED CENTRIFUGE BCBLS-111

LABORATORY LOW SPEED CENTRIFUGE BCBLS-111

BLOOD COLLECTION TUBE DECAPPING CENTRIFUGE



SPECIFICATIONS

Model	BCBLS-111
Max. RCF	5 200 Xg with 4 x 250 ml rotor
Max. Speed	5 000 rpm with 4 x 250 ml rotor
Max. Capacity	4 x 250 ml
Running Time	1 min - 9999 min 59 s, Pulse, Continuous
Drive System	Direct, induction motor drive
Speed Accuracy	± 10 r/min
Programs	12 user-defined
Accel/Decel Profiles	19/19
Noise Level	≤ 60 dB(A)
Power Supply	AC 200-240 V, 50/60 Hz or AC 100-127 V, 50/60 Hz
Max. Power Consumption	600 W
Dimension (W x D x H)	17.7 x 22 x 16.3 in / 45 x 56 x 41.5 cm
Weight w/o Rotor	110 lb / 50 kg
Alt Name	Blood Collection Tube Decapping Centrifuge

ACCESSORIES FOR PURCHASE

No	Name	Rotor Capacity (places x volume, ml)	Tube Size (Φ x L in mm)	Max Speed (rpm)	Max RCF (g)
1	No. 1 Swing-Bucket Rotor	4 x 250 ml	Φ63 x 111	5 000 rpm	5 200 Xg
2	No. 2 Swing-Bucket Rotor	56 x 1.6-7 ml blood collection tube	Φ13 x 75-100	4 000 rpm	3 400 Xg
3	No. 3 Swing-Bucket Rotor	80 x 1.6-7 ml blood collection tube	Φ13 x 75-100	4 000 rpm	3 400 Xg



1



3



Biolab Scientific Ltd.

Trillium Executive Center, East Tower, 675 Cochrane Dr, Markham, Ontario L3R 0B8, Canada
Email: info@biolabscientific.com | Website: www.biolabscientific.com