



## MINI DRY BATH INCUBATOR BDMI-114

# MINI DRY BATH INCUBATOR BDMI-114

## MINI-DRY BATH INCUBATOR



Working area dimension: 3 individual controllers.  
 LCD display. It is easy to set and use.  
 Simultaneously display the set and actual time, temperature, auto-stop and sound-alert after the program completed.  
 Temperature deflection adjusting meets different demands.

## SPECIFICATIONS

Model	BDMI-114
Working area	Block-A / Block-B / Block-C
Temp. control range	0 °C to 99.9 °C / RT. +5 °C to 99.9 °C / RT. +5 °C to 99.9 °C
Heating time (25 °C to 99.9 °C)	≤20 min
Cooling time (20 °C to 0 °C)	≤25 min * Block-A
Temp. control accuracy	≤0.5 °C
Temp. stability @37 °C	0.2 °C
Display accuracy	0.1 °C
Timer range	1~999 h or 1~999 min or 1~999 sec
Block quantity	3
Power supply	DC 12 V, 120 W
Dimension (mm)	360x180x150 mm
Weight	3 kg
* Tested in AT. &lt;25 °C	
Note	It has cooling / heating functions and its temperature ranges from 0 °C~99.9 °C.
Alt Name	Mini-Dry Bath Incubator

## ACCESSORIES FOR PURCHASE

No	Name	Description
1	Block	40x0.2 mL tube
2	Block	24x0.5 mL tube
3	Block	15x1.5 mL tube
4	Block	15x2.0 mL tube
5	Block	8x12.5x12.5 mL cuvette block
6	Block	4x15 mL tube
7	Block	2x50 mL tube
8	Block	



1



2



3



4



5



6



7



**Biolab Scientific Ltd.**

Trillium Executive Center, East Tower, 675 Cochrane Dr, Markham, Ontario L3R 0B8, Canada  
Email: [info@biolabscientific.com](mailto:info@biolabscientific.com) | Website: [www.biolabscientific.com](http://www.biolabscientific.com)