BioLAB

PRODUCT CATALOG



AUTOMATIC GLYCOHEMOGLOBIN ANALYZER BEM-41201

CE ISO9001 CERTIFIED

www.biolabscientific.com

GLYCOHEMOGLOBIN ANALYZER BEM-41201

AUTOMATED GLYCOHEMOGLOBIN ANALYZER



1.Fast
First test takes 2 minutes
1 Min/Test (60 tests/Hour)
Rotated sample rack loader
2.Precise
Ion exchange chromatography / Three elution buffers
Report in both NGSP and IFCC results
3.Stable
4.Easy
One key operation, smart settings
Automatically sampling, hemolyzing and mixing
Applying in whole blood and diluted blood

SPECIFICATIONS

Model	BEM-41201
Measurement Principle	Ion-exchange HPLC
Test Mode	Standard mode/Variant mode
Throughput	Standard mode: 60s/test; Variant mode: 96s/test
Items	Standard mode: FP, A1A, A1B, F, LA1C+, SA1C, A0
Measurement Range	HbA1c: 3%~20%
Repeatability	CV≤1.0%
Sample Type	Whole blood and diluted blood
Sample Management	LIS loading, barcode scanning
Sample Loading Capacity	10 samples/rack*9 racks. Continuous circulating rack sampling system
STAT	STAT priority
HbA1c Report Form	In both graphic and value report with IFCC(mmol/mol) and NGSP(%) results
Consumable	Elution buffer, hemolysis & wash solution, column, filter, calibrator, control
Reagent Management	RFID scanning, automatically loading info of lot, expiry date and remains, etc.
Operation Interface	English operation software, HDMI touch screen
User Management	Authorization set for different levels
Data Storage	100,000 results. Available for statistics, online queries, data backup, data export
Export&Communication	RS-232 (Bi-directional). Built-in thermal printer
Maintenance	Automatic maintenance in startup and off, alarm and error indication, daily maintenance reminder & guidance
Power Supply	AC100~240V, 50/60Hz
Product Size(WDH)	560590490mm
Net Weight	45kg
Packing Size(WDH)	6001100600mm
Gross Weight	58kg
Alt Name	Automated Glycohemoglobin Analyzer

FEATURES

1.Fast
First test takes 2 minutes
1 Min/Test (60 tests/Hour)

Rotated sample rack loader

2.Precise
Ion exchange chromatography / Three elution buffers
Report in both NGSP and IFCC results

3.Stable
4.Easy
One key operation, smart settings
Automatically sampling, hemolyzing and mixing
Applying in whole blood and diluted blood
Mixing loading for primary tubes and micro-cup
RFID scanner, management for reagent lot No., expiry date, and residue check
LIS bi-directional transmission with reports in both diagram and text

Biolab Scientific Ltd.
Trillium Executive Center, East Tower, 675 Cochrane Dr, Markham, Ontario L3R 0B8, Canada
Email: info@biolabscientific.com | Website: www.biolabscientific.com