



CLASS II BIOSAFETY CABINET BER-3804

CLASS II BIOSAFETY CABINET BER-3804

CLASS II A2 BIOSAFETY CABINET



1. Color 7.0 inch touch screen controller
2. Standard with high precision thermal wind sensor
3. With reservation timing sterilization function, three-level authority management function, high efficiency filter, sterilization lamp service life display, built-in socket instrument controllable switch function, front window status display function, can set three position interlock function, cumulative work timing function.
4. Optional: ULPA Filter

SPECIFICATIONS

Model	BER-3804
Classification	A2: 30% efflux, 70% internal circulation
Function	
Cleanliness	Class 100
Collection Efficiency	0.3um particle ≥99.995%
Colony count	≤0.5/vessel (petri dish is Φ90mm)
Noise Level	≤67db
Vibration Semi-peak Value	≤5um
Downflow Velocity	0.33 ± 0.025m/s
Inflow Velocity	0.53 ± 0.025m/s
Lumination	Average luminance ≥650lx, the measured value of each illuminance ≥430lx
Total air exhaust of system	591m³/h
Composition	
Work surface	Work surface
Outer Shell	Outer Shell
Operation Door	Operation Door
HEPA filter (mm)	1120*420*69 / 1580*495*69
Fluorescent Lamp Power	16W*2 (LED)
UV Light Power	40W*1
Fan	860W one piece
Rated Power	2200W
Working height of front window	Working height of front window
Maximum height of front window	440mm
Controller	
Display	Full color 7.0-inch touch screen
Additional Function	Reservation timing sterilization function, three-level authority management function, high efficiency filter, sterilization lamp service life display, built-in socket controller to control switch function, front window status display function, three position interlock function, cumulative work timing function
Specification	
Operating Area Size (W*L*Hmm)	1550*625*628

Exterior Size (W*L*Hmm)	1700*795*2250
Packing Size (W*L*Hmm)	1830*925*1655
Max Power of Universal Socket	Single socket: 800W
NW/GW kg	368/428
Alt Name	CLASS II A2 Biosafety Cabinet



APPLICATIONS

Table type clean equipment.



Biolab Scientific Ltd.

Trillium Executive Center, East Tower, 675 Cochrane Dr, Markham, Ontario L3R 0B8, Canada

Email: info@biolabscientific.com | Website: www.biolabscientific.com