



## GENERAL PURPOSE INCUBATOR BER1P3

# GENERAL PURPOSE INCUBATOR BER1P3

## CONSTANT-TEMPERATURE INCUBATOR



1. New panel 5 degree angle, new rotary switch for glass inner door, convenient to use.
2. The double door structure adopts high quality glass, easy to observe samples, the door with magnetic strip is easy to be opened and closed, good sealing.

## SPECIFICATIONS

|                                 |                                                                                                                       |
|---------------------------------|-----------------------------------------------------------------------------------------------------------------------|
| Model                           | BER1P3                                                                                                                |
| Cycle Mode                      | Natural Convection                                                                                                    |
| Function                        |                                                                                                                       |
| --Temp. Range                   | RT+5-70°C                                                                                                             |
| --Temp. Resolution Ratio        | 0.1°C                                                                                                                 |
| --Temp. Motion                  | ±0.5°C                                                                                                                |
| --Temp. Uniformity              | ±1.5°C                                                                                                                |
| Structure                       |                                                                                                                       |
| --Inner Chamber                 | High strength galvanized sheet                                                                                        |
| --Outer Shell                   | Cold rolling steel electrostatic spraying exterior                                                                    |
| --Insulation Layer              | High quality foam board                                                                                               |
| --Heater                        | Mica electrothermal film                                                                                              |
| --Power Rating                  | 0.6kW                                                                                                                 |
| --Exhaust Hole                  | Φ35mm top(with function of test hole)                                                                                 |
| Controller                      |                                                                                                                       |
| --Temp. Control Mode            | PID Intelligent                                                                                                       |
| --Temp. Setting Mode            | Touch button setting                                                                                                  |
| --Temp. Display Mode            | Measuring temperature: LED upper row; setting temperature: LED lower row                                              |
| --Timer                         | 0-9999min (with timing wait function)                                                                                 |
| --Operation Function            | Fixed temperature operation, timing function, auto stop.                                                              |
| --Additional Function           | Sensor deviation correction, temperature overshoot self-tuning, internal parameter locking power-off parameter memory |
| --Sensor                        | PT100                                                                                                                 |
| Safety Device                   | Over temperature sound-light alarm                                                                                    |
| Specification                   |                                                                                                                       |
| --Inner Chamber size(W*L*H)(mm) | 500*450*550                                                                                                           |
| --Exterior Size (W*L*H)(mm)     | 675*580*820                                                                                                           |
| --Packing Size (W*L*H)(mm)      | 745*662*990                                                                                                           |
| --Volume                        | 125L                                                                                                                  |
| --Max. Shelf Number             | 13                                                                                                                    |

|                          |                                |
|--------------------------|--------------------------------|
| --Load Per rack          | 15kg                           |
| --Shelf Space            | 35mm                           |
| --Power Rating (50/60HZ) | AC220V/2.3A                    |
| --NW/GW (kg)             | 45/60                          |
| Accessory                |                                |
| --Shelf                  | 2                              |
| Optional accessories     | Shelf                          |
| Alt Name                 | Constant-Temperature Incubator |



## APPLICATIONS

Classical bacterial microorganism incubator.



**Biolab Scientific Ltd.**

Trillium Executive Center, East Tower, 675 Cochrane Dr, Markham, Ontario L3R 0B8, Canada

Email: [info@biolabscientific.com](mailto:info@biolabscientific.com) | Website: [www.biolabscientific.com](http://www.biolabscientific.com)