



LABORATORY LOW SPEED CENTRIFUGE BET-3201

LABORATORY LOW SPEED CENTRIFUGE BET-3201

CAPILLARY CENTRIFUGE



SPECIFICATIONS

| | |
|--------------------------|-----------------------------------------------------------------------------------|
| Model | BET-3201 |
| Max speed | 1,200r/min |
| Max RCF | 13,960xg |
| Speed accuracy | ± 30 r/min (customizable, multiples of 10) |
| Control and drive system | High torque AC brushless motor, microcomputer control |
| Timer range | 1min-99h59min59s, with continuous centrifugation and instantaneous centrifugation |
| Noise | ≤ 60 dB(A) |
| Power supply | AC220V50Hz |
| Total power | 400W |
| Weight | 35kg |
| Dimensions (LxWxH) | 550mmx440mmx330mm |
| Alt Name | Capillary Centrifuge |

ACCESSORIES FOR PURCHASE

| No | Name | Description |
|----|-----------------------------|-------------------------------------------------|
| 1 | Capillary rotor 24 slots | Max RPM: 12,000r/min Max RCF: 13,960xg |
| 2 | Capillary rotor 36 slots | Max RPM: 10,000r/min Max RCF: 12,300xg |



1



2

APPLICATIONS

The capillary centrifuge is used in the measurement of hematocrit values and the centrifugation of trace blood and solution.

MORE INFO

A variety of functional blood centrifuges are customized based on the customer's requirements.



Biolab Scientific Ltd.

Trillium Executive Center, East Tower, 675 Cochrane Dr, Markham, Ontario L3R 0B8, Canada

Email: info@biolabscientific.com | Website: www.biolabscientific.com