



LABORATORY LOW SPEED CENTRIFUGE BET1BG1

LABORATORY LOW SPEED CENTRIFUGE BET1BG1

CAPILLARY CENTRIFUGE



SPECIFICATIONS

Model	BET1BG1
Max speed	1,200r/min
Max RCF	13,960xg
Speed accuracy	±30r/min (customizable, multiples of 10)
Control and drive system	High torque AC brushless motor, microcomputer control
Timer range	1min-99h59min59s, with continuous centrifugation and instantaneous centrifugation
Noise	≤60dB(A)
Power supply	AC220V50Hz
Total power	400W
Weight	35kg
Dimensions (LxWxH)	550mmx440mmx330mm
Alt Name	Capillary Centrifuge

ACCESSORIES FOR PURCHASE

No	Name	Description
1	Capillary rotor 24 slots	Max RPM: 12,000r/min Max RCF: 13,960xg
2	Capillary rotor 36 slots	Max RPM: 10,000r/min Max RCF: 12,300xg



1



2

APPLICATIONS

The capillary centrifuge is used in the measurement of hematocrit values and the centrifugation of trace blood and solution.

MORE INFO

A variety of functional blood centrifuges are customized based on the customer's requirements.



Biolab Scientific Ltd.

Trillium Executive Center, East Tower, 675 Cochrane Dr, Markham, Ontario L3R 0B8, Canada
Email: info@biolabscientific.com | Website: www.biolabscientific.com