



LABORATORY LOW SPEED CENTRIFUGE BFH1CC1 (BCFL-301-B)

LABORATORY LOW SPEED CENTRIFUGE BFH1CC1

LOW SPEED CENTRIFUGE



1. Microprocessor control, less noise, it is widely used to qualitative analysis of blood serum, plasma and urea in the fields of hospital, blood center, laboratory and biochemistry.
2. Brushless Frequency motor, free maintenance, no powder pollution, quick in speed up and down.
3. The flexible axle driven system which drive the rotor directly, smooth in operation and small vibration.
4. Micro computer control system, LED display the RCF,time and speed.
5. compact design, super speed and imbalance protection.

SPECIFICATIONS

Model	BFH1CC1
Old Model	BCFL-301-B
Max. Speed	5000rpm
Max. RCF	4140xg
Max. Capacity	6x100ml
Time Range	0~99min59sec
RPM/RCF Convert	Yes
Noise	≤ 55dBA
Speed Accuracy	±20rpm
Acc/Dec	10 Kinds
Temperature range	Normal
Power supply	AC 220V/110V 50HZ/60HZ
Size (W x D x Hmm)	525x370x280mm
Net Weight(Kg)	28KG
Certificates	CE,ISO & Calibration report are available
Alt Name	Low Speed Centrifuge

ACCESSORIES FOR PURCHASE

No	Name	Max Speed (rpm)	Max Volume (ml)	Max. RCF (xg)
1	Swing rotor	5000	12x10ml	4140
2	Swing rotor	5000	8x15ml	4110
3	Fixed rotor	4000	24x15ml	2240
4	Fixed rotor	5000	12x15ml	3080
5	Fixed rotor	5000	6x100ml	3130
6	Fixed rotor	5000	6x50ml	2850



Biolab Scientific Ltd.

Trillium Executive Center, East Tower, 675 Cochrane Dr, Markham, Ontario L3R 0B8, Canada

Email: info@biolabscientific.com | Website: www.biolabscientific.com