



PH METER BFU-5203

PH METER BFU-5203

PH METER



SPECIFICATIONS

Model	BFU-5203
Parameters	Conductivity
Range Cond	(0~2x10 ⁵) uS/cm
Range TDS	-
Range Temp	-
Accuracy Cond	±1.0%FS ±1bit
Accuracy TDS	-
Accuracy Temp	-
Temp. Compensation	(5 ~ 35)°C, manual
Power	220V/110V
Dimension/N.W.	300x200x90 mm / 1.5kg
Shipping size/G.W.	330x260x280 mm / 2kg
Standard Accessories	Electrode Holder
Alt Name	Conductivity meter



Biolab Scientific Ltd.

Trillium Executive Center, East Tower, 675 Cochrane Dr, Markham, Ontario L3R 0B8, Canada

Email: info@biolabscientific.com | Website: www.biolabscientific.com