



VACUUM OVEN BGI-1104

VACUUM OVEN BGI-1104

TWO&THREE CHAMBERS VACUUM OVEN

LCD controller (with timing function)

Vacuum oven is designed especially for drying of material which is thermo-sensitive or decomposes and oxidative easily. It can be filled with inert gases, which is especially for a rapid drying of some compound material.



- Microprocessor controller.
 - Dual layer tempered glass door for clear observation.
 - Minimum heating time 50% less than traditional vacuum oven.
 - Interior chamber made from 304 stainless steel with a mirror polished finish making for easy cleaning.
 - Standard with vacuum pump.
 - Independent audible and visible temperature-limiting alarm system ensures experiments run safely (Option).
 - Each chamber can be heated, vacuumed, and injected with nitrogen independently without interfering with each other.
- Option:
- LED lamp

SPECIFICATIONS

| | |
|-------------------------------|---|
| Model | BGI-1104 |
| Electrical Requirement | AC220V 50HZ |
| Controller | Controller |
| Power Consumption | 3100W |
| Temperature Range | RT+10 ~ 200°C |
| Display Resolution | 0.1 |
| Temperature Stability | ±1°C |
| Vacuum Degree | 133Pa |
| Ambient Temperature | +5~40°C |
| Chamber Volume | 140Lx3 |
| Interior Dimension (WxDxH) mm | 380x600x620/Chamber |
| Exterior Dimension (WxDxH) mm | 880x980x1910 |
| Shelves | 2pcs /Chamber (Independent temp. control) |
| Independent temp. Controller | 3 PCS |
| Chamber Material | 304 Stainless steel |
| Packing Dimension | 118*100*209 |
| CBM | 2.47 |
| Net Weight | - |
| Gross weight | - |
| Alt Name | Vacuum Oven (3 chambers) |

FEATURES

- Microprocessor controller.
- Dual layer tempered glass door for clear observation.
- Minimum heating time 50% less than traditional vacuum oven.
- Interior chamber made from 304 stainless steel with a mirror polished finish making for easy cleaning.
- Standard with vacuum pump.
- Independent audible and visible temperature-limiting alarm system ensures experiments run safely (Option).

- Each chamber can be heated, vacuumed, and injected with nitrogen independently without interfering with each other.

Option:

- LED lamp
- Inert gas valve
- Oil mist filter



Biolab Scientific Ltd.

Trillium Executive Center, East Tower, 675 Cochrane Dr, Markham, Ontario L3R 0B8, Canada

Email: info@biolabscientific.com | Website: www.biolabscientific.com