



## STABILITY TEST CHAMBER BGI1BH11 (BTST-1611)

# STABILITY TEST CHAMBER BGI1BH11

## LARGE MEDICINE STABILITY TESTING CHAMBER



## SPECIFICATIONS

|                               |   |
|-------------------------------|---|
| Model                         | BGI1BH11  |
| Old Model                     | BTST-1611   |
| Temperature control range     | 0~65°C without Illumination 15~50°C with Illumination   |
| Temperature Stability         | ±0.5°C  |
| Temperature Uniformity        | ±2°C ~ ±3°C   |
| Humidity Range                | 35~95%RH  |
| Humidity Stability            | ±3%RH   |
| Illumination                  | 0~6000LX adjustable   |
| Illumination difference       | ≤±500LX   |
| Cooling systemcooling mode    | Two sets of Imported compressor work rotationally   |
| Controller                    | Programmable (touch screen)   |
| Sensor                        | Temp: Pt100, Humidity: capacitance sensor   |
| Ambient Temperature           | RT+5~30°C   |
| Electrical Requirement        | AC220V 50Hz   |
| Power Consumption             | 3600W   |
| Operating Power               | 2880W   |
| Chamber Volume                | 800L  |
| Interior Dimension (WxDxH,mm) | 800x590x1650  |
| External Dimension (WxDxH,mm) | 1360x890x2000   |
| Shelves                       | 4(pcs)  |
| Safety Device                 | Compressor overheating and overpressure protection, Fan overheating protection, Over temperature protection, Overload protection, Water protections |
| Remark                        | 1. Build in mini printer2. Paperless recorder (Option)3. GSD series products have Installed Intensity of Illumination detector                      |
| Alt Name                      | Large Medicine Stability Testing Chamber  |



**Biolab Scientific Ltd.**

Trillium Executive Center, East Tower, 675 Cochrane Dr, Markham, Ontario L3R 0B8, Canada  
Email: [info@biolabscientific.com](mailto:info@biolabscientific.com) | Website: [www.biolabscientific.com](http://www.biolabscientific.com)