



COOLED INCUBATOR BGI1U9 (BICL-303)

COOLED INCUBATOR BGI1U9

COOLING INCUBATOR



Microprocessor control

- Polished stainless-steel chamber, semicircular arcs at corners for easy cleaning, and the space between the shelves in the chamber is adjustable.
- Latest PID controller
- R134a refrigerant, imported compressor
- Independent temperature-limiting alarm system ensures experiments run safely.(Option)
- Printer connector and RS485 connector are options which can connect printer and computer to record the parameters and the variations of temperature.(Option)
- There is a 25mm instruction connection hole on the left side of the chamber for easy testing operation and temperature measurement.

SPECIFICATIONS

Model	BGI1U9
Old Model	BICL-303
Temperature Range	20°C~65°C
Display Resolution	0.1°C
Temperature Stability	High Temp $\pm 0.5^{\circ}\text{C}$ / Low Temp $\pm 1.0^{\circ}\text{C}$
Electrical Requirement	220V 50Hz
Ambient Temperature	+5°C~35°C
Power Consumption	1600W
Interior Dimension (WxDxH)mm	600x500x830
Exterior Dimension (WxDxH)mm	740x900x1580
Shelves	3(pcs)
Timing Range	0~9999min
Alt Name	Cooling Incubator

FEATURES

Microprocessor control

- Polished stainless-steel chamber, semicircular arcs at corners for easy cleaning, and the space between the shelves in the chamber is adjustable.
- Latest PID controller
- R134a refrigerant, imported compressor
- Independent temperature-limiting alarm system ensures experiments run safely.(Option)
- Printer connector and RS485 connector are options which can connect printer and computer to record the parameters and the variations of temperature.(Option)
- There is a 25mm instruction connection hole on the left side of the chamber for easy testing operation and temperature measurement.

Option:

- Printer
- Independent temperature-limiting Alarm system
- RS485 connector

APPLICATIONS

Serve for preserve culture medium, serum, medicine as well as microorganism training and environmental testing etc.



Biolab Scientific Ltd.

Trillium Executive Center, East Tower, 675 Cochrane Dr, Markham, Ontario L3R 0B8, Canada
Email: info@biolabscientific.com | Website: www.biolabscientific.com