



ORBITAL SHAKER BGI1Z6 (BSOT-607)

ORBITAL SHAKER BGI1Z6


ORBITAL SHAKER



- Microprocessor PID controller, with timing function.
- Stainless-steel or alnico plate, non-eroding and easy clean.
- DC Brushless motor with long using life, wide speed control with long using life, wide speed control.
- The alarm will get off when the real shaking speed is 10% different with setting value, and motor will stop automatically.
- Speed controller ensures smooth start /stop which can prevent the liquid spill to damage the equipment.
- Microprocessor PID controller, with timing function.

SPECIFICATIONS

Model	BGI1Z6
Old Model	BSOT-607
Electrical Requirement	220V 50Hz
Type of Movement	Orbital
Shaking Speed Range	40~250 rpm
Amplitude	20mm
Temperature Range	RT+5~100°C
Display Resolution	0.1°C
Power Consumption	400W
Platform Size (mm)	400x340
Default Platform	Spring wire racks
Alt Name	Orbital Shaker



FEATURES

- Microprocessor PID controller, with timing function.
 - Stainless-steel or alnico plate, non-eroding and easy clean.
 - DC Brushless motor with long using life, wide speed control with long using life, wide speed control.
 - The alarm will get off when the real shaking speed is 10% different with setting value, and motor will stop automatically.
 - Speed controller ensures smooth start /stop which can prevent the liquid spill to damage the equipment.
-
- Microprocessor PID controller, with timing function.
 - Stainless-steel or alnico plate, non-eroding and easy clean.
 - DC Brushless motor with long using life, wide speed control with long using life, wide speed control.
 - The alarm will get off when the real shaking speed is 10% different with setting value, and motor will stop automatically.
 - Speed controller ensures smooth start /stop which can prevent the liquid spill to

damage the equipment.

- They are with heating plate function.

APPLICATIONS

Microprocessor controller (with timing function)

Widely applicable for researches on bacteria cultivation, fermentation, hybridization, chemical and biochemical reaction.



Biolab Scientific Ltd.

Trillium Executive Center, East Tower, 675 Cochrane Dr, Markham, Ontario L3R 0B8, Canada

Email: info@biolabscientific.com | Website: www.biolabscientific.com