



SCANNING UV VISIBLE SPECTROPHOTOMETER BHV-501

SCANNING UV VISIBLE SPECTROPHOTOMETER BHV-501

SCANNING SPECTROPHOTOMETER

It is an advanced single beam design, it provides excellent performance for measurements.

It is suitable for clinical, pharmaceutical, and bio-chemical lab applications, as well as routine applications such as quantitative analysis, kinetics, Wavelength Scan, Multi-Wavelength and DNA/Protein analysis. The memory is 32K.

UV-Vis Analyst software Based Microsoft Windows makes these instruments versatile. All instruments provide excellent performance for measurements.

PC models

1. Online software update via the internet.
2. Data can be downloaded.
3. The PC models come standard with Windows based application software UV-Vis Analyst.



Fixed bandwidths.

Sealed, solvent-resistant tactile keypad with alpha-numeric entry for file names and units.

Pre-aligned deuterium lamp for easy lamp replacement. The status of the lamps may be monitored

PC Windows based software UV/Vis Analyst including sophisticated utility programs.

Data Download-to-PC software for stand-alone models (Optional)

Real-time clock for date and time stamping of results.

Data can be saved by a USB memory device directly.

USB memory port

Brief Function Introduction:

SPECIFICATIONS

Model	BHV-501
Wavelength Range	190-1100 nm
Spectral Bandwidth	2 nm
Optical System	Single Beam Grating 1200 lines/mm
Wavelength Accuracy	±0.5 nm
Wavelength Repeatability	0.3 nm
Scan Speed	Hi, MED, LOW, MAX. 3000 nm/min
Photometric Accuracy	±0.5%T or ±0.004A@1A
Photometric Range	0-200%T, -0.3-3A, 0-9999Conc
Noise	≤0.0001A @ 0.0A (500 nm, RMS)
Stray Light	0.05%T @ 220 nm, 340 nm
Stability	0.0008A/h @ 500 nm
Display	5 inches LCD (320x240 dots)
Baseline Flatness	±0.002A
Standard Cell Holder	Standard 10 mm pathlength cuvette
Light Source	Tungsten & Deuterium lamp
Output	USB Type A (USB memory, right side); USB Type B (PC connectivity, back); Parallel port for printer
Power Requirement	AC 110/220V 50/60Hz
Dimension (L*W*H)	491x365x180 mm
Weight	14 kg
Alt Name	Scanning Spectrophotometer

FEATURES

Fixed bandwidths.

Sealed, solvent-resistant tactile keypad with alpha-numeric entry for file names and units.

Pre-aligned deuterium lamp for easy lamp replacement. The status of the lamps may be monitored

PC Windows based software UV/Vis Analyst including sophisticated utility programs.


Data Download-to-PC software for stand-alone models (Optional)

Real-time clock for date and time stamping of results.

Data can be saved by a USB memory device directly.




USB memory port



Brief Function Introduction:


All methods are included as built-in standard; this eliminates the need for software. Online software update via the internet.

The local control software includes functions such as: Photometry, Quantitative, Wavelength Scan, Kinetics, DNA/Protein, Multi-wavelength and System Utilities.




Standard Curve:

Up to 10 standard solutions may be used to establish a calibration equation curve. There is a choice of four methods for fitting a curve through the calibration points: Linear fit, Linear fit through zero, square fit and cubic fit.




Multi-Wavelength:

Up to 10 wavelengths may be entered, allowing the measurement of multiple wavelengths on a series of samples.



Wavelength Scan:

The Wavelength Scan intervals are 0.1, 0.2, 0.5, 1, 2, 5nm, and High, Medium and Low scan speeds are available. Scan speeds vary from 100 to 2000 nm/min. Wavelengths are scanned from high to low so that the instrument stands by at high wavelength. This minimizes the degradation of UV sensitive samples. Precise control of filter and lamp changes means that their effects are not seen on the final scan. Post-run manipulation includes re-scaling axes, curve tracking and peak picking.



Kinetics:

This mode may be used for scanning time courses or reacting rate calculations. Abs. vs. time graphs displayed on the screen in real time.

Wait time and measurement time up to 12 hours may be entered with time intervals of 0.5, 1, 2, 5, 10, 30, seconds and 1 min.

Post-run manipulation includes re-scaling, curve tracking and selection of the part of the curve required for the rate calculation. Rate is calculated using a linear regression algorithm before multiplying by the entered factor.

Item	Result	Unit
A1	0.251	Abs
A2	0.243	Abs
Pref	0.095	Abs
C-DNA	4.521	ug/ul
C-Pro	110.8	ug/ul
Ratio	1.059	

DNA/Protein Test:

Concentration and DNA purity are calculated by

Absorbance ratios 260nm/280nm or 260nm/230nm with optional subtracted absorbance at 320nm

DNA Concentration= $62.9 \cdot A_{260} - 36.0 \cdot A_{280}$

Or $49.1 \cdot A_{260} - 3.48 \cdot A_{230}$

Protein Concentration= $1552 \cdot A_{260} - 757.3 \cdot A_{280}$

Or $183 \cdot A_{260} - 75.8 \cdot A_{230}$

Other wavelengths and factors may be entered.



Biolab Scientific Ltd.

Trillium Executive Center, East Tower, 675 Cochrane Dr, Markham, Ontario L3R 0B8, Canada

Email: info@biolabscientific.com | Website: www.biolabscientific.com