



GLOSS METER BJH1Q1 (BMET-905)

GLOSS METER BJH1Q1

GLOSS METER

Accurate Gloss Meter is with independent intellectual property, manufactured according to international standard ISO2813 and China standard GB/T 9754. With autocalibration and highend QC software, it can meet the first grade requirements of JJG696. The gloss meter also features a USB/RS232 interface and Bluetooth 2.1 dual mode.



Elegant Design combined with aesthetics and ergonomics;
With autocalibration function;
Meet the first grade requirements of JJG696;
Large storage to save over 35000 data;
Auto Poweroff within optional 30-120 seconds;
Comply to ISO2813, ASTM523, GB/T9754;
With PC software for quality report, more extend functions;
With multi working modes and multifunctions for most customers' requirements.

SPECIFICATIONS

Model	BJH1Q1
Old Model	BMET-905
Measuring Angle	60°
Measuring Range	0-200 GU
Measuring Area	9x15 mm
Division Value	1 GU
Measuring Modes	Basic mode
Measuring Time	0.5 s
Repeatability	0-100 GU: ± 0.5 GU; 100-2000 GU: $\pm 0.5\%$ GU
Accurate	Conform with JJG696 second grade requirements of gloss meter
Auto Power-off Time	30 s
Language	Chinese & English
Display	2.3 inch white and black screen
Size	160 x 52 x 84 mm
Weight	About 300 g
Power Supply	1 pc dry-cell battery (can measure 10000 times) or USB charge
Interface	USB
Operation Temperature Range	0-40 °C (32-104 °F)
Storage Temperature Range	-20-50 °C (-4-122 °F)
Humidity Range	< 85% RH, without condensation
Standard Accessories	User manual, calibration plate
Optional Accessories	Miniature printer
Notification	The technical parameter is only for reference
Alt Name	Gloss Meter



APPLICATIONS

Economic Gloss Meter, apply to paint, ink, stoving varnish, coating, wood products; marble, granite, vitrified polished tile, pottery brick and porcelain; plastic, paper; hardware industries, etc. for gloss measurement and gloss data transmission.



Automobile



Leather



Plastic



Metal



Stone Material



Coating



Biolab Scientific Ltd.

Trillium Executive Center, East Tower, 675 Cochrane Dr, Markham, Ontario L3R 0B8, Canada
Email: info@biolabscientific.com | Website: www.biolabscientific.com