



## DRYING OVEN BODR-7003

# DRYING OVEN BODR-7003

## DRYING OVEN



LCD Programmable controller (with timing function)  
Polished stainless-steel chamber, semicircular arcs at corners for easy cleaning, and the space between the shelves in the chamber is adjustable.  
Auto-controller of fan speed to prevent damage to the samples.  
Large LCD screen to display more data at same time.  
Self-check function makes it easy to identify problems.  
There is a 25mm instruction connection hole on the left side of the chamber for easy testing operation and temperature.  
Programmable controller: 7 periods 63 steps, 0~5999min for each period , fan speed 0 to 100% adjustable.  
Independent audible and visible temperature-limiting alarm system ensures experiments run safely.

## SPECIFICATIONS

Model	BODR-7003
Electrical Requirement	220V 50HZ
Controller	Programmable LCD display
Power Consumption	1550W
Temperature Range	RT+10~250°C
Display Resolution	0.1°C
Temperature Stability	±1°C
Temperature Uniformity	±2.5% (@100°C)
Shelves	3(PCS)
Interior Dimension (WxDxH)mm	500x380x750
Exterior Dimension (WxDxH)mm	655x715x980
Chamber Volume	150L
Timing Range	0~5999min
Packing Dimension	83*73*115
CBM	0.69
Net Weight	81.00
Gross weight	101
Alt Name	Drying Oven

## FEATURES

LCD Programmable controller (with timing function)  
Polished stainless-steel chamber, semicircular arcs at corners for easy cleaning, and the space between the shelves in the chamber is adjustable.  
Auto-controller of fan speed to prevent damage to the samples.  
Large LCD screen to display more data at same time.  
Self-check function makes it easy to identify problems.  
There is a 25mm instruction connection hole on the left side of the chamber for easy testing operation and temperature.  
Programmable controller: 7 periods 63 steps, 0~5999min for each period , fan speed 0 to 100% adjustable.  
Independent audible and visible temperature-limiting alarm system ensures experiments run safely.  
RS485 connector can connect computer and printer to record the parameters and the variations of temperature.(Option)



**Biolab Scientific Ltd.**

Trillium Executive Center, East Tower, 675 Cochrane Dr, Markham, Ontario L3R 0B8, Canada  
Email: [info@biolabscientific.com](mailto:info@biolabscientific.com) | Website: [www.biolabscientific.com](http://www.biolabscientific.com)