



TESTER PETROLEUM EQUIPMENT BPTL-403

TESTER PETROLEUM EQUIPMENT BPTL-403

ULTRAVIOLET FLUORESCENCE SULFURIN-OIL ANALYZER

The instrument is utilized to decide the whole sulfur substance by bright fluorescence method. It improves the capacity of anti-jamming and avoids the complicated operation of titration pool and components of insecurity which utilized Coulometry.



The instrument is used to determine the total sulfur content by ultraviolet fluorescence method. It Improves the ability of anti-jamming and avoids the complicated operation of titration pool and factors of instability which used Coulometry. So the sensitivity of the instrument is greatly improved. The data collecting, processing, storage and printing are fully controlled by computer.

SPECIFICATIONS

Model	BPTL-403
Sample injection quantity	Solid: 1-20 mg; Liquid: 5-20 μ L; Gas: 1-5 mL
Determination method	Ultraviolet fluorescence method (S)
Measuring range	5 ppm ~ 5000 ppm (High concentration should be diluted, low concentration gas sample is up to 0.1 ppm)
Temperature range	Ambient to 1150°C
Temperature control precision	$\pm 1^\circ\text{C}$
Air supply - High purity argon	Above 99.9%
Air supply - High purity oxygen	Above 99.9%
Power supply	AC220V $\pm 22\text{V}$, 50Hz $\pm 0.5\text{Hz}$, 1500 W
Dimension - Host	305(W) x 460(D) x 440(H) mm
Dimension - Temp controller	550(W) x 460(D) x 440(H) mm
Net weight - Host	20 kg
Net weight - Temp controller	40 kg
Standard configuration	Printer + Computer + SYD 0689 + Liquid injector
Other optional parts	Solid sample injector, gas sample injector
Alt Name	Ultraviolet Fluorescence Sulfurin-Oil Analyzer

APPLICATIONS

Petroleum Industry, Oil and Gas Industry



Biolab Scientific Ltd.

Trillium Executive Center, East Tower, 675 Cochrane Dr, Markham, Ontario L3R 0B8, Canada

Email: info@biolabscientific.com | Website: www.biolabscientific.com