



HEATING MANTLE

BER-4813 BER-4814 BER-4815 BER-4816 BER-4817 BER-4818

INDEX

Preface	2
Warranty	2
1. Safety Instructions	4
2. Proper Uses	5
3. Inspections	5
3.1 Receiving Inspection	6
3.2 Listing of Items	6
4. Trial Runs	7
5. Operating Modes	8
6. Faults	8
7. Maintenance and Cleaning	9
8. Storage and transportation	10
9. Working condition	10

Preface

Thank you for purchasing our products: Heating Mantle. Users should read this Manual carefully, follow the instructions and procedures, and beware of all the cautions when using this instrument.

Warranty

This instrument is warranted to be free from defects in materials and workmanship under normal use and service, for a period of 12 months from the date of invoice. The warranty is extended only to the original purchaser. It shall not apply to any product or parts which have been damaged on account of improper installation, improper connections, misuse, accident or abnormal conditions of operation. For claims under the warranty please contact us. You may also send the instrument directly to our works or we will send you the spare parts to help you resolve this problem in the next order, enclosing the invoice copy and by giving reasons for the claim. You would be solely liable for freight cost.



Connect the device to an earthed power supply to ensure safety of machine and experiment; connect the power as the machine required.



This equipment is forbid to use in inflammable and explosive, poisonous and strong corrosive experiments.



Make sure horizontal installation.



Non-professionals are not allowed to disassemble and repair this machine.



Pay attention to the set temperature while dealing with the inflammable matters.



Make sure dry the resin container, if the temperature is setting too high by accident, the container would be dissolved and then fall on the heater to cause fire.



Overfilling of sample will lead to overheat of working room under parts, which will dissolve the inflammable material and cause fire.



While the machine is working, don't touch the top, window and exhaust port of the device to protect from high-temperature burns.



Read the instruction book before operation.

Table 1

1. Safety Instructions

- When working, wear the personal guard to avoid the risk from:
 - Splashing and evaporation of liquids
 - Release of toxic or combustible gases.
- Set up the instrument in a spacious area on a stable, clean, non-slip, dry and fireproof surface; do not operate the instrument in explosive atmospheres, with hazardous substances or under water.
- Temperature must always be set to at least 25°C lower than the fire point of the media used.
- Beware of hazards due to:
 - Flammable material or media with a low boiling temperature
 - Overfilling of media
 - Unsafe container
- Process pathogenic materials only in closed vessels.
- Check the instrument and accessories before hand for damage each time you use them. Do not use damaged components. Safe operation is only guaranteed with the accessories described in the "Accessories" chapter. Accessories must be securely attached to the device and cannot come off by themselves. Always disconnect the plug before fitting accessories.
- Ensure that the external temperature sensor is inserted in the media to a depth of at least 20mm.
- When using metal vessels, do not place the temperature sensors on the bottom of the vessel. Placing sensors on the vessel bottom can cause excessively high temperature to be measured especially in media which have poor conductivity. The tip of the measuring sensor must be at least 5mm from the vessel bottom, a distance of 10mm is ideal.
- The instrument can only be disconnected from the main power supply by pulling out the mains plug or the connector plug.
- The voltage stated on the label must correspond to the main power supply.
- Ensure that the mains power supply cable does not touch the heating base plate. Do not cover the device.
- Keep away from high magnetic field.

2. Proper Uses

The instrument is designed for mixing and / or heating liquids in schools, laboratories or factories. This device is not suitable for using in residential areas.

3. Inspections

3.1 Receiving Inspection

Unpack the equipment carefully and check for any damages which may have arisen during transport. If it happens, please contact manufacturer for technical support.

Note:

If there is any apparent damage to the system,
Please do not plug it into the power line.

3.2 Listing of Items

The packing includes the following items:

Magnetic Heating Mantle

Items Qty

Main Unit 1

Power Cable 1

User Manual 1

Stir Bar 1

Please check the instrument and appendix with the packing list when you first open the instrument packing case.

If you find there is something wrong with the instrument and the appendix, do contact the vendor or the producer.

4. Trial Runs

- Make sure the required operating voltage and power supply voltage match.
- Ensure the socket must be earthed reliably.
- Ensure the power be off
- Plug in the power cable, ensure the power be on and begin initializing.
- Add the medium into the vessel with a stirring bar if with the magnetic stirrer function.
- Put the vessel on the work plate.
- Adjust the stirring speed and start stirring if with the magnetic stirrer function..
- Observe the stirring bar and LCD display if with digital function.
- Adjust the temperature and start heating.
- Observe the real temperature on LCD display if with the digital function.
- Stop the heating and stirring functions.

If these operations above are normal, the device is ready to operate. If these operations are not normal, the device may be damaged during transportation, please contact manufacture for technical support.

5. Operating Modes

Rows Manual Heating Mantle:

- Place the equipment on level worktable, and then put the container with liquid inside into heating mantle.
- Switch on the power accords with the machine, then power indicator light will be lighten; turn on the power of regulation knob, and turn the knob clockwise, then the working indicator light will be lighten; in the process of regulating, the light intensity changes according to different regulation position, and the temperature rises as well.
- The same to adjust the magnetic stirring power.



Notes:

The power must accord with the machine. Adjust the speed slowly; please adjust the speed when too high speed makes the stir breakaway. Make sure the power line has safety distance from the



heating mantle.

When the machine meets fault please cut off the Electricity first.

6. Faults

- Instruments can't be power ON

Check whether the power cable is plugged

Check whether the fuse is broken or loose

- Temperature cannot reach set point or stirring can't be starts when adjust the control knob

Check whether the heating wire broke during transport

Check whether the controller broke during transport

If these faults are not resolved, please set the instruments to factory default setting, or take the unit to your technical service center, or contact with the manufacturer.

7. Maintenance and Cleaning

- Proper maintenance can keep instruments working in a good state and lengthen its lifetime.
- Be careful not spray the cleanser into the instrument when cleaning.
- Unplug the power line when cleaning.
- Only use cleanser that we advised as below:

Dyes Isopropyl alcohol

Construction materials Water containing tenside/

Isopropyl alcohol

Cosmetics water containing tenside/

Isopropyl alcohol

Foodstuffs Water containing tenside

Fuels Water containing tenside

8. Storage and transportation

- Keep it in dry and clean room with good ventilation and no corrosive gas
- prevent it from wetting by the rain and avoid violent collision in transportation.

9. Working condition

- Ambient temperature: 5~40°C;

- Ambient humidity: ≤90%;

Voltage: 220V±10%, 50/60Hz or 110V+/-10%, 50/60Hz

Model	Capacity (ml)	Voltage (V)	Max Temp	Power (W)	Working Time	Exterior Size (mm)	Packing Size (mm)	N.W (KG)	G.W (KG)
BER-4813	100	220V, 50/60Hz and 110V/60Hz	450°C	100x2	Continuous	280x140x106	320x180x190	4	5
BER-4814	250	220V, 50/60Hz and 110V/60Hz	450°C	150x2	Continuous	320x160x170	360x190x200	4	5
BER-4815	500	220V, 50/60Hz and 110V/60Hz	450°C	250x2	Continuous	360x180x180	400x220x210	4	5
BER-4816	1000	220V, 50/60Hz and 110V/60Hz	450°C	350x2	Continuous	420x210x200	460x250x230	7	8

BER-4817	100	220V, 50/60Hz and 110V/60Hz	450°C	100x4	Continuous	530x140x160	570x180x190	8	10
BER-4818	250	220V, 50/60Hz and 110V/60Hz	450°C	150x4	Continuous	610x160x170	650x190x200	8	10

Table 2



Biolab Scientific Ltd.

Trillium Executive Center, East Tower, 675 Cochrane Dr, Markham, Ontario L3R 0B8, Canada
Email: info@biolabscientific.com | Website: www.biolabscientific.com